



The client logo displayed on this certificate is reproduced with permission. The presence of the logo does not confer UKAS accreditation on that organisation and the client is in no way associated with UKAS or Eurofins other than as a customer of Eurofins.

Client: Pozzani Pure Water plc
Warwick Road
Fairfield Industrial Estate
Louth
Lincs
LN11 0YB
United Kingdom

Certificate No.: 2932300
Job Ref: 19A04245
Number Of Samples: 1
Date Received: 09/01/2019
Date Reported: 23/01/2019

CERTIFICATE OF ANALYSIS

Page 1 of 2

LSN: 2R/16725
Sample Description.: 61430
Sample Taken: 07/01/2019
Testing Commenced: 09/01/2019

Cust. Details:
Site Details:

Site details printed here.

Test Description	Method	Sample Condition:		Acceptable SPEC	Grade
		Result	Unit		
† Total Arsenic	Subcontracted	13	µg/l As	10	R
† Total Copper	Subcontracted	3.4	µg/l Cu	2000	G
‡ Total Iron	Subcontracted	7100	µg/l Fe	200	R
† Total Lead	Subcontracted	0.5	µg/l Pb	10	G
† Total Manganese	Subcontracted	210	µg/l Mn	50	R
† Total Zinc	Subcontracted	4	µg/l Zn	3000	G
Total Aerobic Colony Count 37°C 2 Da MW3		None Detected	cfu/ml	20	G
Total Aerobic Colony Count 22°C 3 Da MW3		4	cfu/ml	100	G
Total Coliforms	MW17	None Detected	MPN/100ml	1	G
Escherichia coli	MW17	None Detected	MPN/100ml	1	G
† Total Chloride	Subcontracted	24	mg/l Cl	250	G
† Total Sulphate	Subcontracted	200	mg/l SO4	250	G
† Fluoride	Subcontracted	0.18	mg/l F	1.5	G
† Nitrate	Subcontracted	<0.5	mg/l NO3	49.6	G
† Nitrite	Subcontracted	<0.1	mg/l NO2	0.5	G
† Ammonia	subcontracted	0.27	mg/l N	0.39	G
† pH	Subcontracted	7.31	pH value	6.5-9.5	G
‡ Electrical Conductivity	Subcontracted	850	µS/cm	2500	G
‡ Total Hardness	Subcontracted	540	mg/l CaCO3		
‡ Turbidity	Subcontracted	49	NTU	4	R
‡ Odour (Qualitative)	Subcontracted	Not Detected			

Results presented above are a measure of the quality of the water sample received (as is) by Eurofins. These results may not be a true measure of the quality of the sample at point of origin if prescribed sampling and transportation protocols have not been adhered to.
The lower limit of detection (LOD) for legionella is 25cfu per volume or swab tested, apart from the centrifuge method for which there is an LOD of 3cfu/150ml.

Signed for and on behalf of Eurofins Water Hygiene Testing UK Limited

Claire Brechin
MRSC Rsci
Laboratory Team Leader



9658

Eurofins Water Hygiene Testing UK Ltd,
i54 Business Park,
Valiant Way,
Wolverhampton,
WV9 5GB,
UK.

T: +44(0)1506 534 317
E: info@eurofins.co.uk
W: www.eurofins.co.uk



The client logo displayed on this certificate is reproduced with permission. The presence of the logo does not confer UKAS accreditation on that organisation and the client is in no way associated with UKAS or Eurofins other than as a customer of Eurofins.

Client: Pozzani Pure Water plc
Warwick Road
Fairfield Industrial Estate
Louth
Lincs
LN11 0YB
United Kingdom

Certificate No.: 2932300
Job Ref: 19A04245
Number Of Samples: 1
Date Received: 09/01/2019
Date Reported: 23/01/2019

CERTIFICATE OF ANALYSIS

Page 2 of 2

†	Total Calcium	Subcontracted	170	mg/l Ca	500	G
†	Total Magnesium	Subcontracted	28	mg/l Mg		
†	Total Sodium	Subcontracted	15	mg/l Na	200	G

Results presented above are a measure of the quality of the water sample received (as is) by Eurofins. These results may not be a true measure of the quality of the sample at point of origin if prescribed sampling and transportation protocols have not been adhered to.
The lower limit of detection (LOD) for legionella is 25cfu per volume or swab tested, apart from the centrifuge method for which there is an LOD of 3cfu/150ml.

Signed for and on behalf of Eurofins Water Hygiene Testing UK Limited

Claire Brechin
MRSC Rsci
Laboratory Team Leader



Testing Laboratory – Eurofins Livingston, 6-8 Cochrane Square, Brucefield Industry Park, Livingston, EH54 9DR
Registered Office: i54 Business Park, Valiant Way, Wolverhampton, WV9 5GB, UK. Registered No.9805380



9658